

Instructions for Universal Alkaline Phosphatase Immunostaining Kit (For Mouse and Rabbit Primary Antibodies) Cat. # KA-50FM

Intended Use:

This product is intended for qualitative immunohistochemistry with normal and neoplastic formalin-fixed, paraffin-embedded tissue sections, to be viewed by light microscopy. Clinical interpretation of staining results should be accompanied by histological studies with proper controls. Patients' clinical histories and other relevant diagnostic tests should be utilized by a qualified person (s) when evaluating and interpreting results.

This kit is designed to label specific primary antibodies immunohistochemically on tissue sections. The immunohistochemical protocol defined in this brochure is just a guideline. We encourage the individual laboratory to optimize its own protocol based upon tissue fixation conditions, primary antibodies employed, and the user's experience. These reagents were tested and quality controlled using tissue sections, however, they can also be optimized for cell smears and cytospin preparations. Sufficient reagents are provided to run 50 to 100 tests.

Test principle:

The high affinity of the non covalent interaction between biotin and streptavidin (1×10^{15}) forms the basis for this immunostaining kit. It requires the formation of an irreversible and specific linkage between biological macromolecules. The immunohistochemical applications of the interaction between avidin and biotin were introduced by Bayer et al. (1979), who described techniques for generating active biotinyl compounds such as biotin-N-hydroxysuccinimide and biotin hydrazine and for conjugating them to various organic compounds, including immunoglobulins and alkaline phosphatase (ALP). Streptavidin (SA) is a tetrameric protein (mol. wt. $4 \times 15,000$), isolated from the actinobacterium *Streptomyces avidinii* (Chalet & Wolf, 1964). Streptavidin can bind to four molecules of biotin. Streptavidin gives superior results compared to avidin because its isoelectric point is closer to neutral pH, whereas avidin is positively charged at physiological pH. Streptavidin does not carry any carbohydrate side chain, whereas avidin is composed of 70% of carbohydrate. Because of this, SA does not have the tendency to bind non-specifically. Primary antibodies bind to the target antigens in the tissue sections and conjugated secondary antibodies bind specifically to these receptor antibodies. A biotin conjugated secondary antibody, in turn, is traced by a streptavidin conjugated enzyme and can be visualized by an appropriate substrate/chromogen.

Introduction:

Immunohistochemical techniques are spreading rapidly and the practice of anatomic pathology has undergone a revolutionary change since the development of these procedures (Nadji & Morales, 1983). Because of their versatility, sensitivity, and specificity, immunoperoxidase stains are invariably the best stains when appropriate antibodies are available. With the ever increasing number of antibodies against cellular antigens, immunoperoxidase techniques now provide a powerful tool to resolve a wide array of diagnostic pathology. All immunohistochemical techniques require the specific antibody employed to be labeled that they can be easily seen when attached to cellular antigens. At the same time the sensitivity of the immunoperoxidase techniques are central to wide variety of specific antigen localization. Alkaline phosphatase based kits are of special value for staining tissues that have high endogenous peroxidase activity. Alkaline phosphatase is unaffected by endogenous peroxidase and therefore results in cleaner background. Our kit is based on direct SA-ALP conjugate technology. The linker reagent is biotinylated anti-mouse immunoglobulins, capable of labeling mouse primary antibodies.

Reagents Supplied:

IVD: For In Vitro Diagnostic Use

DBS will not be held responsible for patent infringement or other violation that may occur with the use of our product

DBS

1020 Serpentine Lane, # 114, Pleasanton, CA 94566 Tel: 925 484 3350, Fax: 925 484 3390

Website: www.dbiosys.com e-mail: customersupport@dbiosys.com

- Bottle 1 Alkaline Phosphatase Enhancer: 10mL clear solution.
Used to increase the intensity of signal.
- Bottle 2 Linker Reagent: 10mL clear yellow solution of biotinylated anti-mouse
immunoglobulins.
- Bottle 3 ALP Tracer Reagent: 10mL clear cherry red solution of conjugated streptavidin alkaline
phosphatase.
- Detector Reagents: Alkaline Phosphatase Buffer: 2.5mL clear solution.
Fast Red TR chromogen: 10 concave white/beige tablets
Naphthol ASM-MX Substrate: 0.5mL clear solution.

Precautions:

- i) Several reagents in this kit contain sodium azide. Follow instruction provided by your local authorities for disposal. If disposed in the sink, flush the drain pipe to avoid a reaction of sodium azide in the plumbing system.
- ii) Once you start the immunostaining process, don't let tissue sections dry because it can cause undesirable background staining and artifacts.
- iii) Interpretation of the results will be the sole responsibility of the user.

Storage: All of the kit components should be stored at 2-8°C. Do not freeze. Do not use beyond the expiration date stated on the label.

Material Required but not provided:

Xylene or dewaxing agents
Absolute alcohol
Distilled or de ionized water
Counter stain
Immuno Wash buffer (DBS catalog # K 005)
Mounting medium (DBS catalog # K 002)
Oven or incubator
Staining jars
Microscope slide
Microscope
Diamond pencil
Absorbent pads
Primary antibodies

Specimen Preparation:

The purpose of immunohistochemistry is to localize, identify, and (whenever possible) quantitate cell and tissue constituents. Optimal fixation of the tissue is the key to good immunostaining. Successful immunostaining depends on the accurate preservation of the epitopes looked for in its structural context. We recommend our users read an excellent article on tissue preparation by Larsson (1993) to understand the need and importance of optimal tissue fixation.

IVD: For In Vitro Diagnostic Use

DBS will not be held responsible for patent infringement or other violation that may occur with the use of our product

DBS

1020 Serpentine Lane, # 114, Pleasanton, CA 94566 Tel: 925 484 3350, Fax: 925 484 3390

Website: www.dbiosys.com e-mail: customersupport@dbiosys.com

Reagent Preparation: Except for the chromogen, all reagents are provided in ready to use form.

Preparation of Substrate/Chromogen:

- i) Transfer 5mL of de ionized or distilled water in a test tube.
- ii) Add five drops of concentrated Alkaline Phosphatase Buffer and mix.
- iii) Add one tablet of Fast Red to the water/Alkaline Phosphatase Buffer mixture. Let it sit for a minute and then shake or vortex to dissolve the tablet.
- iv) Add one drop of Naphthol AS-MX Substrate and mix.

The chromogen/substrate solution is stable for at least 1 hour. Any solution not used during this period should be discarded.

Positive and Negative controls:

Each immunostaining run should include a known positive and a negative control to assure the proper functioning of staining system and valid interpretation of the results.

Positive control:

A tissue which is known to contain the desired antigen and has given positive staining in the past.

Negative control: One of the following should be used as negative control:

- i) Instead of primary antibody, use the normal nonimmune serum from the same species of animal in which the primary antibody was raised.
- ii) Use buffer in which the primary antibody was diluted.
- iii) Use a tissue known not to contain the desired antigen.
- iv) Absorb the primary antibody with the appropriate antigen and use it instead of primary antibody.

Staining Protocol:

Step I Removal of paraffin wax: Deparaffinize tissue sections according to the established procedure in your lab and bring tissues to wash buffer.

Step II Endogenous Alkaline Phosphatase Blocking: This is an optional step and should be performed only if tissue is suspected to have high endogenous alkaline phosphatase activity. Apply enough drops of Alkaline Phosphatase Blocker to cover tissue. Incubate for 5-10 minutes at room temperature.

Step III Washing: Drain off excess reagent. Rinse with wash buffer three times for 1 minute each time. Drain off excess buffer and carefully wipe slide around the tissue to remove excess buffer from the glass leaving the tissue wet.

Immunostaining Protocol I:

This protocol is recommended for optimally fixed tissues with abundant antigens in the tissue and for high affinity primary antibodies.

Step I Primary antibody: Apply enough drops of primary antibody to cover the tissue section. Incubate according to the manufacture's recommended conditions. Wash and wipe slides as described above.

IVD: For In Vitro Diagnostic Use

DBS will not be held responsible for patent infringement or other violation that may occur with the use of our product

DBS

1020 Serpentine Lane, # 114, Pleasanton, CA 94566 Tel: 925 484 3350, Fax: 925 484 3390

Website: www.dbiosys.com e-mail: customersupport@dbiosys.com

- Step II Linker reagent: Apply enough drops of linker reagent to cover the tissue section. Incubate for 10 minutes at room temperature. Wash slides as described previously.
- Step III Tracer Reagent: Apply enough drops of tracer reagent to cover the tissue section. Incubate for 10 minutes at room temperature. Wash slides as described previously.
- Step IV Substrate/Chromogen: Apply working chromogen solution for 5-10 minutes at room temperature for color development. For best results, look under the microscope for signal development. Once desired signal to noise ratio is achieved, stop the reaction by washing slides in wash buffer. Note development time and follow it during subsequent incubations.

Immunostaining Protocol II:

This protocol is recommended for less than optimally fixed tissues with low antigenic density in the tissue and for low affinity primary antibodies.

- Step I Primary antibody: Apply enough drops of primary antibody to cover the tissue section. Incubate according to the manufacture's recommended conditions. Wash slides as described previously.
- Step II Linker reagent: Apply enough drops of linker reagent to cover the tissue section. Incubate for 20 minutes at room temperature. Wash slides as described previously.
- Step III Tracer Reagent: Apply enough drops of tracer reagent to cover the tissue section. Incubate for 20 minutes at room temperature. Wash slides as described previously.
- Step IV Substrate/Chromogen: Apply working chromogen solution for 5-10 minutes at room temperature for color development. For the best results, look under the microscope for the signal development. Once desired signal to noise ratio is achieved, stop the reaction by washing the slide in wash buffer. Note development time and follow it during subsequent incubations.
- Step V Wash slides and counter stain with an appropriate counter stain. Mount and observe staining under the microscope.

Enhanced Staining:

If additional signal is desired, treat tissue sections with Alkaline Phosphatase Enhancer for 1 minute after Step III. Drain off excess enhancer and add substrate/chromogen solution without any wash.

- References:**
- (i) Bayer et al., *Method Enzymol.* 62, 308 (1979).
 - (ii) Chaiet & Wolf, *Arch. Biochem. Biophys.* 106, 1 (1964).
 - (iii) Giorno, *Diag. Immuno.* 2, 161 (1984).
 - (iv) Hsu et al., *J Histo. Cyto.* 29, 577 (1981).
 - (v) Nadji & Morales, *Anat. pathol.* 14, 767 (1983).
 - (vi) Nagle et al. *J Histo. Chem.* 31, 1010 (1983).
 - (vii) Larsson, *Applied Immunohistochem.* 1, 2, (1993).
 - (viii) Petrusz & Ordronneau, In Polak & Van noorden eds., *Immunocytochemistry: Practical applications in pathology and biology.* Bristol wright-PSG, 212, (1983).

IVD: For In Vitro Diagnostic Use

DBS will not be held responsible for patent infringement or other violation that may occur with the use of our product

DBS

1020 Serpentine Lane, # 114, Pleasanton, CA 94566 Tel: 925 484 3350, Fax: 925 484 3390

Website: www.dbiosys.com e-mail: customersupport@dbiosys.com