

Montage™ Hematoxylin

Catalog Number: K056-100AN

Document #: DS-2040-B
Effective Date: 12/18/2023

Intended Use

For In Vitro Diagnostic Use.

Product Description

Hematoxylin is a histological staining reagent suitable for visualization of nuclei in tissue sections. Montage™ Hematoxylin, is designed for optimal performance when used together with the Montage 360™ System.

Summary and Explanation

DBS's Hematoxylin is intended for use in the histologic demonstration of nuclear staining. This staining technique is used to make the critical distinction between a normal nucleus and an abnormal one.

Format

Ready To Use

Volume/UOM

100 mL

Storage and Handling

Store at room temperature. Do not use after expiration date printed on label.

Preparation of Working Solutions

1. Montage Hematoxylin is ready to use and does not require any preparation.

Protocol Recommendations

1. Deparaffinize and Re-hydrate tissue with water
2. Apply enough Hematoxylin to completely cover the tissue section and incubate for 2-10 minutes with Montage Hematoxylin according to desired intensity
3. Rinse in Tap or DI water for 20 – 30 seconds
4. Incubate with bluing reagent for 30 seconds
5. Wash with Tap or DI water.
6. Follow suggested dehydration and mounting procedures based on Chromogen used.

Quality Control

Refer to CLSI Quality Standards for Design and Implementation of Immunohistochemistry Assays; Approved Guideline-Second edition (I/LA28-A2). CLSI Wayne, PA, USA (www.clsi.org). 2011.

Troubleshooting

Contact Diagnostic BioSystems Technical Support at (925) 484-3350, extension 2, techsupport@dbiosys.com or your local distributor to report unusual staining results.

Warranty

There are no warranties, expressed or implied, which extend beyond this description. Diagnostic BioSystems is not liable for property damage,

personal injury, or economic loss caused by this product.

Performance Characteristics

The protocols for a specific application can vary. These include, but are not limited to: fixation, heat-retrieval method, incubation times, tissue section thickness and detection kit used. Due to the superior sensitivity of these unique reagents, the recommended incubation times and titers listed are not applicable to other detection systems, as results may vary. The data sheet recommendations and protocols are based on exclusive use of Diagnostic BioSystems products. Ultimately, it is the responsibility of the investigator to determine optimal conditions. These products are tools that can be used for interpretation of morphological findings in conjunction with other diagnostic tests and pertinent clinical data by a qualified pathologist.

Precautions

This product is a single-use, non-sterile, in vitro diagnostic device.

1. Wear disposable gloves when handling reagents.
2. Specimens, before and after fixation, and all materials exposed to them should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come in contact with sensitive areas, wash with copious amounts of water.
3. Microbial contamination of reagents may result in an increase in nonspecific staining.
4. Incubation times or temperatures other than those specified may give erroneous results. The user must validate any such change.
5. Do not use reagent after the expiration date printed on the label.
6. The MSDS is available upon request.
7. Consult OSHA, federal, state or local regulations for disposal of any toxic substances.

