

Oil Red O Stain Kit (For Fat)

Description: Oil Red O Stain Kit (For Fat) is intended for use in the histological visualization of fat cells and neutral fat. This kit may be used **ONLY** on frozen tissue sections, fresh smears, or touch preps.

Fat Cells:	Red
Neutral Fat:	Red
Nuclei:	Blue

Uses/Limitations: For In-Vitro Diagnostic use only.
Histological applications.
Do not use past expiration date.

Control Tissue: Any frozen section containing fat.

Availability/Contents:

<u>Kit Contents</u>	<u>Volume</u>	<u>Storage</u>
Propylene Glycol	500 ml	15-30°C
Oil Red O Solution	125 ml	15-30°C
Hematoxylin, Mayer's (Lillie's Mod.)	125 ml	15-30°C

Precautions: This product is a single-use, non-sterile, in vitro diagnostic device.
Keep away from open flame.
Avoid contact with skin and eyes.
Harmful if swallowed.
Follow all Federal, State, and local regulations regarding disposal.
Use in chemical fume hood whenever possible.

Procedure (Standard):

1. Prepare fresh or frozen tissue section as usual.
 2. Place slide in Propylene Glycol for 2 minutes.
 3. Incubate slide in Oil Red O Solution for 6 minutes.
- Note: Prepare mixture of 85% Propylene Glycol in distilled water.
4. Differentiate tissue section in 85% Propylene Glycol for 1 minute.
 5. Rinse slide in 2 changes of distilled water.
 6. Incubate slide in Hematoxylin, Mayer's (Lillie's Modification) for 1-2 minutes.
 7. Rinse slide thoroughly in tap water.
 8. Rinse slide in 2 changes of distilled water.

Storage: 15° C



30° C

**Store All Components At
Room Temperature.**



Diagnostic BioSystems
6616 Owens Drive
Pleasanton, CA 94588
U.S.A.
Tel: 1-925-484-3350

c V

P

EmergoEurope
Prinsessegracht 20
2514 AP The Hague
The Netherlands
Tel: (+31)(0) 70 345-8570

Instructions For Use
KT025-IFU

Rev. Date: June 12, 2018

Revision: 5


Page 2 of 2

6616 Owens Drive, Pleasanton, CA 94588 U.S.A. - Tel. (925) 484-3350 - Fax (925) 484-3390 - www.dbiosys.com

9. Coverslip using an aqueous mounting medium (cat# AML060).

References:

1. Sheehan, DC., Hrapchak, BB. Theory and Practice of Histotechnology; 1980, page 225.

Storage: 15° C  30° C

**Store All Components At
Room Temperature.**



Diagnostic BioSystems
6616 Owens Drive
Pleasanton, CA 94588
U.S.A.
Tel: 1-925-484-3350

C V

P

EmergoEurope
Prinsessegracht 20
2514 AP The Hague
The Netherlands
Tel: (+31)(0) 70 345-8570